


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130184 ✓
Name	JOHN HAMILTON	ID#	3192 Date 6/2/09 ✓
A	Agency	PETERSBURG POLICE DEPARTMENT	Phone # 772-3838
	Instrument Location	PETERSBURG ✓	
Alco S/N		X 39600 ✓	Target Value .085 ✓ High Pressure 500 PSI
B	Alco Test Values	.086 ✓ 1 st Alco	.086 ✓ 2 nd Alco
Signature		 ✓	OB 6/5/09
(OVER)			

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.


Nita J. Bolz

Scientific Director

State Breath Alcohol Program


Subscribed and sworn before me this 24th day of June 2009.


Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office

BT9 03/00



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130184 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130184 ✓

JUNE 02, 2009 ✓

OPERATOR'S NAME:
HAMILTON JOHN A
OPERATOR'S NUMBER: 3192
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
A

O.L. #: A
DEPT AGENCY: PS61
CASE REPORT: A
TEST TYPE: U
ALCO TARGET VALUE: .085 ✓
ALCO S/N: K301600 ✓

--- BREATH ANALYSIS ---

.085 ADJUSTED FOR 30.32 in		
ALCO TARGET	.086	05:56
BLANK TEST	.000	05:57
INTERNAL STANDARD	VERIFIED	05:57
ALCO TO 30.32 in	.085 ✓	05:58
BLANK TEST	.000	05:58
SUBJECT SAMPLE	.000	05:59
BLANK TEST	.000	06:00
ALCO TO 30.33 in	.086 ✓	06:00
BLANK TEST	.000	06:01

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130184

JUNE 02, 2009 ✓
TIME 06:02

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	30.33 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

"#5%(")*+,-./0123456789:;<=>?@AB CDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefg h i j k l m n o
p q r s t u v w x y z { | } ~